



SpecCheck Secondary Gel-Standards Kit for DPD Chlorine Methods

Product #:	2980500
ZAR Price:	Contact Hach
Available	

Chlorine SpecCheck Secondary Standards Kit for Mid Range DPD Chlorine Methods. Contains a blank vial and 3 colored vials at ~0.2, ~1.6, and ~2.7 mg/L Cl_2 . Ready-to-use SpecCheck Secondary Standards Kits include four 10 mL (1 inch) sealed glass cells with a blank and three different standards within the typical testing range, instructions, plastic case and Certificate of Analysis. SpecCheck Standards are computer-formulated and have a two-year shelf life from date of manufacture.

Note: For use only with Mid Range DPD chlorine methods (Pocket Colorimeter II, DR300 and DR900).

Not for use with Low Range methods (including LR/HR Pocket Colorimeter II).

*This product has not been evaluated to test for chlorine and chloramines in medical applications in the United States.

Solid Gel Standards - No preparation required

Ideal for instrument verification in the lab and in the field

2-year guaranteed performance

Specifications

Concentration:	0 - 4.0 mg/L Cl_2
Method:	DPD
Parameter:	Chlorine, DPD MR
Quantity:	Set of 4 Vials