



Cobalt / Nickel Reagent Set

Product #: 2242600
ZAR Price: Contact Hach
Ships within 3 days

For the determination of Cobalt, 0.01-2.00 mg/L Co, and Nickel, 0.006-1.000 mg/L Ni, PAN Method. For 25 mL sample.

Specifications

Description:	Nickel Reagent Set, PAN Method, 25 mL
Instrument:	DR/3000, DR/2000, DR900
Method:	8078
Method Number 2:	8150
Number of tests:	100
Parameter:	Cobalt, Nickel
Range:	0.01 - 2.00 mg/L Co
Range 2:	0.006 - 1.000 mg/L Ni
Sample Volume:	25 mL
Special Features:	Cobalt Nickel Reagent set, 25ml

Required Accessories

- Gloves, Disposable, Powder Free, Nitrile, Large (Item 2550503)