



## Reagent set, carbon dioxide, digital titrator

**Product #:**

**2272700**

**ZAR Price:**

Contact Hach

**Ships within 2 weeks**

 **Hazardous**

Items with this mark may be considered hazardous under some shipping conditions.

If necessary, we will change your selected shipping method to accommodate these items.

For determination of Carbon Dioxide by Digital Titrator titration. Hach Method 8205. Range: 10 - 1000 mg/L as CO<sub>2</sub> Set includes: Phenolphthalein Pillows and two Sodium Hydroxide Digital Titrator Cartridges (0.3636 N and 3.3636 N). Approximately 40-100 tests

### Accurate results - when and where you need them

Digital Titrator methods use a high-precision dispensation device and cartridges of concentrated titrant to allow analyses in the lab, the plant or the field.

### Fast and Flexible

Interchangeable cartridges allow you to change titrants in seconds. No cleaning or rinsing burets.

### Versatile

More than 50 parameter/range combinations are available

### Economical

Concentrated titrants allow 40-100 typical titrations per cartridge.

---

## Specifications

Method:	8205
Number of tests:	40 - 100
Platform:	Digital Titrator
Range:	10 - 1000 mg/L CO <sub>2</sub>

---

## Required Accessories

- Gloves, Disposable, Powder Free, Nitrile, Large (Item 2550503)