



Chlorine, total, Reagent Powder Pillows, 100/pk, 10 mL

Product #: 2105669
ZAR Price: Contact Hach
Available

For total chlorine determination by the DPD Method.

USEPA accepted for reporting for drinking and wastewater analysis using Hach Method 8167.

Range: 0.02 - 2.00 mg/L Cl_2 . Sample Size: 10 mL.

Pack of 100 powder pillows.

Boric Acid free!

Specifications

EPA compliant:	Yes
Footnote:	*Generally, the DR 5000 performs the same methods as the DR/4000. The DR 2800 performs the same methods as the DR 2700, DR/2500, and DR/2400. Some exceptions do exist.
Instrument:	All except DR/3000, DR/2000
Method:	8167
Method Number 2:	8031
Method Number 3:	8016
Number of tests:	100
Parameter:	Total Chlorine, Bromine, Iodine
Platform:	Powder Pillows
Range:	0.02 - 2.00 mg/L Cl_2
Range 2:	0.05 - 4.50 mg/L Br_2
Range 3:	0.07 - 7.00 mg/L I_2
Reference method:	SM 4500-Cl G
Sample size:	10 mL
Storage conditions:	10 - 25 °C

Required Accessories

- Gloves, Disposable, Powder Free, Nitrile, Large (Item 2550503)