

DPD Total Chlorine Reagent Powder Pillows, 25 mL, pk/100

Product #: **1406499**

ZAR Price: Contact Hach
Available

For total chlorine determination by the DPD method. USEPA accepted for reporting drinking and wastewater analysis using Hach Method 8167.

Ranges: Low Range: up to - 2.00 mg/L; Mid Range: 0.05-4.00 mg/L; High Range: 0.1-10.0 mg/L.

Sample Size: Low Range - 25 mL; Mid Range: 10 mL; High Range: 5 mL.

Pack of 100 powder pillows.

Boric Acid free!

Specifications

EPA compliant:	Yes
Instrument:	DR/3000, DR/2000
Method:	DPD
Method Number 2:	10070 HR
Method Number 3:	10250 MR
Number of tests:	100
Parameter:	Total Chlorine, Bromine, Iodine
Platform:	Powder Pillows
Range:	0.02 - 2.00 mg/L Cl_2
Range 2:	0.1 - 10.0 mg/L Cl_2
Range 3:	0.05 - 4.00 mg/L Cl_2
Sample size:	25 mL
Storage conditions:	10 - 25 °C

Required Accessories

- Gloves, Disposable, Powder Free, Nitrile, Large (Item 2550503)