



## Silicone Wiper for TU5300sc and TU5400sc Automatic Cleaning Module

**Product #:**

**LZY915**

**ZAR Price:**

Contact Hach

No availability shown

Maintain a clean vial with the replacement silicone wiper for the Automatic Cleaning Module. The wiper is recommended for replacement every six months, or as needed depending on sample type. For samples containing high levels of iron and manganese, order part number LZQ176.

LZY915 is the replacement item for LZQ165.